

GB Photometer DEHA (N, N-Diethylhydroxilamin)

● Operation



Switch the unit on using the ON/OFF switch.

dh

The display shows the following:

Fill a clean vial with the water sample up to the 10 ml mark, replace the cap tightly and place the vial in the sample chamber making sure that the Δ-mark on the vial aligned with the ∇-mark on the instrument.



Press the ZERO/TEST key.



The method symbol flashes for approx. 3 seconds.

0.0.0

The display shows the following:

After zero calibration is completed, remove the vial from the sample chamber.
Add the appropriate reagent tablet / solution ; a colour will develop in the sample.
Replace the cap tightly and place the vial in the sample chamber with the Δ and ∇ marks aligned.



Press the ZERO/TEST key.



The method symbol flashes for approx. 3 seconds.

RESULT

The result appears in the display.

Repeating the analysis:

Press the ZERO/TEST key again.

New zero calibration:

Press the MODE key until the desired method symbol appears in the display again.

● User messages

EOI

Light absorption too great. Reasons: zero calibration not carried out or, possibly, dirty optics.

+Err

Measuring range exceeded or excessive turbidity.

-Err

Result below the lowest limit of the measuring range.

LO BAT

Replace 9 V battery, no further analysis possible.

● Technical data

Light source:	LED: $\lambda = 528 \text{ nm}$ (filter)
Battery:	9 V block battery (life = approx. 600 tests)
Auto-OFF:	Automatic switch off 15 minutes after last keypress
Ambient conditions:	5-40°C 30-90% rel. humidity (non-condensing).
CE:	DIN EN 55 022, 61 000-4-2, 61 000-4-8, 50 082-2, 50 081-1, DIN V ENV 50 140, 50 204

● DEHA 0.02-0.5 mg/l

0.0.0

Perform zero calibration (see "Operation").

Hold the drip bottle vertically and add evenly sized 6 drops to the 10 ml water sample by pressing slowly. Replace the cap tightly and invert the vial gently to mix the contents.

Add one DEHA tablet straight from the foil to the same sample and crush using a clean stirring rod.

Mix well using the stirring rod to dissolve the tablet.

Replace the cap tightly and place the vial in the sample chamber with the Δ and ∇ marks aligned.

Wait for a colour reaction time (in darkness) of 10 minutes!



Press the ZERO/TEST key.

dh

The method symbol flashes for approx. 3 seconds.

RESULT

The result is shown in the display in mg/l DEHA.

Measuring tolerance: $\pm 0.05 \text{ mg/l DEHA}$

● Note

Always adhere to the sequence of reagent addition.

● Method notes

Observe application options, analysis regulations and matrix effects of methods. Reagent tablets / solutions are designed for use in chemical analysis only and should be kept well out of the reach of children.

If necessary, request safety data sheets.


Ensure proper disposal of reagent solutions.

● Factors for the below listed oxygen scavengers

(multiply the displayed value with the appropriate factor)

Hydroquinone	5
Erythorbic Acid	7
Methylethylketoxime	7

● Correct filling of the vial



● Troubleshooting: Guidelines for photometric measurements

1. Vials, caps and stirring rods should be cleaned thoroughly **after each analysis** to prevent errors being carried over. Even minor reagent residues can cause errors in the test results. Use the brush provided for cleaning.
2. The outside of the vial must be clean and dry before starting the analysis. Clean the outside of the vials with a towel. Fingerprints or other marks will be removed.
3. Zero calibration and test must be carried out with the same vial as there may be slight differences in optical performance between vials.
4. The vials must be positioned in the sample chamber for zero calibration and test with the Δ-mark on the vial aligned with the ∇-mark on the instrument.
5. Zero calibration and test must be carried out with the sample chamber lid closed.
6. Bubbles on the inside of the vial may also lead to errors. In this case, fit the vial with a clean stopper and remove bubbles by swirling the contents before starting test.
7. Avoid spillage of water or reagent solution in the sample chamber. If water should leak into the photometer housing, it can damage electronic components and cause corrosion.
8. Contamination of the windows over the light source and photo sensor in the sample chamber can result in errors. If this is suspected check the condition of the windows.
9. The reagent tablets should be added to the water sample without being handled.
10. Large temperature differentials between the photometer and the operating environment can lead to incorrect measurement due to the formation of condensate in the area of the lens or on the vial (e.g).
11. Method specified at $20 \text{ °C} \pm 2 \text{ °C}$.
12. To avoid errors caused by stray-light do not use the instrument in bright sunlight.

● Calibration mode



Press MODE key and **keep it depressed**.



Switch unit on using ON/OFF key.
Release MODE key after approx. 1 second.

CAL

The display shows the following in alternating mode:

dh



Perform zero calibration (see "Operation").
Press the ZERO/TEST key.



The method symbol flashes for approx. 3 seconds.

0.0.0

The display shows the following in alternating mode:

CAL



Place the calibration standard to be used in the sample chamber with the Δ and ∇ marks aligned.
Press the ZERO/TEST key.



The method symbol flashes for approx. 3 seconds.

RESULT

CAL

The result is shown in the display, alternating with CAL.

If the result displayed corresponds with the value of the calibration standard (within the tolerance quoted), exit calibration mode by pressing the ON/OFF key.



Otherwise, pressing the MODE key once increases the displayed value by 1 digit.



Pressing the ZERO/TEST key once decreases the displayed value by 1 digit.

CAL

Pressing the relevant key until the displayed value equals the value of the calibration standard.

RESULT + x



By pressing the ON/OFF key, the new correction factor is calculated and stored in the user calibration software.

: : Confirmation of calibration (3 seconds).

● Note

CAL

Factory calibration active.

cAL

Calibration has been set by the user.

● Recommended calibration value

DEHA: between 0.2 and 0.4 mg/l DEHA

● User calibration : cAL

Manufacturing calibration : CAL

To reset the calibration to the factory setting:



Press both the MODE and ZERO/TEST and **keep them depressed**.



Switch the unit on using the ON/OFF key. Release the MODE and ZERO/TEST keys after approx. 1 second.

The following messages will appear in turn on the display:

SEL

The calibration is reset to the factory setting.

CAL

(SEL stands for Select)

or:

SEL

Calibration has been set by the user. (If the user calibration is to be retained, switch the unit off using the ON/OFF key.)

cAL



Calibration is reset to the factory setting by pressing the MODE key. The following messages will appear in turn on the display:

SEL

CAL



Switch the unit off using the ON/OFF key.

● User notes

E 10

Calibration factor "out of range"

E 70

Manufacturing calibration incorrect / erase

E 71

User calibration incorrect / erase